


声 明
Statement

1. 本报告无检验检测专用章、报告骑缝章和批准人签章无效。
This report is invalid without special seal of inspection, cross-page seal and the approver's signatures.
2. 本报告页面所使用“PONY”、“谱尼”字样为本单位的注册商标, 其受《中华人民共和国商标法》保护, 任何未经本单位授权的擅自使用和仿冒、伪造、变造“PONY”、“谱尼”商标均为违法侵权行为, 本单位将依法追究其法律责任。
The pattern and character “PONY” and “谱尼” are registered trademarks of this unit, which are protected by the Trademark Law of the People's Republic of China. Any unauthorized use, imitation, forgery or alteration of the “PONY” and “谱尼” trademarks is an illegal infringement, and this unit will依法追究 its legal responsibility.


古医学实验室: (027) 83997121	武汉实验室: (027) 83997127	北京医学实验室: (010) 62457723-8010	黑龙江医学实验室: (0451) 58427755
太原实验室: (0351) 7555721	武汉医学实验室: (027) 85446975	北京潘尼科技公司: (010) 80415661	郑州实验室: (0371) 65
成都实验室: (028) 8770278	杭州实验室: (0571) 87219096	杭州潘尼医学实验室: (0571) 88706866	郑州潘尼医学实验室: (0371) 65
常州实验室: (0512) 85221000	杭州医学实验室: (0571) 87219096	青岛实验室: (0532) 88706866	杨建实验室: (0532) 88706866
苏州实验室: (0512) 62997900	南京实验室: (0771) 5518818	大连医学实验室: (0411) 87336618	
呼和浩特实验室: (0471) 450025	武汉医学实验室: (027) 85446975		

检测日期				检测环境	
(Test Date)				(Test Environment)	
2021-12-14~2021-12-23				符合要求	
检测项目		样品状态		见附表	
(Test Items)		见下页		褐色固体	

检测结论		样品名称		见附表	
(Test Conclusion)		见下页		见附表	



桃绿色东方12月 份无机化合物	钡	mg/L	<25	3.28	符合		
	镍	mg/L	<0.5	<0.02	符合		
	砷	mg/L	<0.3	0.102	符合		
	总铬	mg/L	<4.5	0.32	符合		
	六价铬	mg/L	<1.5	<0.004	符合		
	硒	mg/L	<0.1	0.00459	符合		
	含水率	%	<30	18.8	符合		



检测报告

(Test Report)

No. B2BC13034001

第 3 页, 共 4 页 (page 3 of 4)

附表 1. 检测项目方法仪器

检测项目 (Test Items)	分析方法 (Test Methods)	仪器和设备 (Instrument and Equipment)	检出限 (Detection Limits)
砷	固体样品 砷、硒、碲 电感耦合等离子体-微波消解 原子荧光法 HJ 702-2014	原子荧光光谱仪	0.00002mg/L

仪器设备 (Instrument and Equipment)	仪器型号 (Instrument Model)	仪器编号 (Instrument Number)
电子分析天平	ML204	B2-IE229-25
紫外分光光度计	UV-1800	B2-IE006-07
电感耦合等离子体发射光谱仪	5110 ICP-OES	B2-IE001-05

